

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022620\
 Data File : VX015014.D
 Acq On : 26 Feb 2020 12:44
 Operator : JC/SP
 Sample : L1683-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-118.5T-20200225

Quant Time: Feb 26 13:45:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	427834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	678486	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	653106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	335437	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	243655	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
35) Dibromofluoromethane	5.49	113	205912	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	8.71	98	824612	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	11.13	95	303226	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	118596	24.519	ug/l	90
44) Trichloroethene	7.21	130	30327	5.761	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022620\
Data File : VX015014.D
Acq On : 26 Feb 2020 12:44
Operator : JC/SP
Sample : L1683-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-118.5T-20200225

Quant Time: Feb 26 13:45:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

