

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022620\
 Data File : VX015021.D
 Acq On : 26 Feb 2020 15:22
 Operator : JC/SP
 Sample : L1683-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140N-20200225

Quant Time: Feb 27 03:15:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	418064	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	657182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	636509	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	326836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	239881	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	
35) Dibromofluoromethane	5.49	113	200989	49.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.50%	
50) Toluene-d8	8.71	98	792821	51.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.82%	
62) 4-Bromofluorobenzene	11.14	95	298625	53.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.50%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.19	85	3247	0.820	ug/l	93
4) Vinyl Chloride	1.41	62	3668	0.759	ug/l #	88
9) 1,1,2-Trichlorotrifluoroet	2.38	101	2665	0.680	ug/l #	89
12) 1,1-Dichloroethene	2.37	96	8704	2.271	ug/l	97
21) trans-1,2-Dichloroethene	3.16	96	10245	2.423	ug/l	95
27) cis-1,2-Dichloroethene	4.59	96	306425	64.831	ug/l	90
44) Trichloroethene	7.21	130	1550160	304.016	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022620\
Data File : VX015021.D
Acq On : 26 Feb 2020 15:22
Operator : JC/SP
Sample : L1683-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140N-20200225

Quant Time: Feb 27 03:15:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

