

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022620\
 Data File : VX015027.D
 Acq On : 26 Feb 2020 17:38
 Operator : JC/SP
 Sample : L1683-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPRING-20200225

Quant Time: Feb 27 03:36:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	403991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	641963	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	628141	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	320792	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	231593	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
35) Dibromofluoromethane	5.50	113	196532	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	8.71	98	783585	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
62) 4-Bromofluorobenzene	11.13	95	290725	53.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.14%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	4042	0.885	ug/l	85
44) Trichloroethene	7.21	130	28678	5.758	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022620\
Data File : VX015027.D
Acq On : 26 Feb 2020 17:38
Operator : JC/SP
Sample : L1683-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPRING-20200225

Quant Time: Feb 27 03:36:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

