

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
 Data File : VX040456.D
 Acq On : 26 Feb 2024 13:33
 Operator : JC/MD
 Sample : VX0226MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0226MBL01

Quant Time: Feb 27 05:07:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	149862	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73377	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67940	46.993	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.980%
35) Dibromofluoromethane	5.379	113	50661	48.613	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.220%
50) Toluene-d8	8.646	98	185519	49.043	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.079	95	72107	50.006	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
Data File : VX040456.D
Acq On : 26 Feb 2024 13:33
Operator : JC/MD
Sample : VX0226MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0226MBL01

Quant Time: Feb 27 05:07:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

