

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
 Data File : VX040465.D
 Acq On : 26 Feb 2024 17:02
 Operator : JC/MD
 Sample : P1578-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20240222

Quant Time: Feb 27 05:11:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	127758	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59980	50.295	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.580%	
35) Dibromofluoromethane	5.385	113	43137	48.555	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.100%	
50) Toluene-d8	8.647	98	161943	50.217	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	11.079	95	63405	51.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.160%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	1421	1.601	ug/l	84
9) 1,1,2-Trichlorotrifluo...	2.331	101	18634	23.029	ug/l	99
12) 1,1-Dichloroethene	2.319	96	7197	8.859	ug/l	92
16) Acetone	2.380	43	1200	3.085	ug/l	99
20) Methylene Chloride	2.788	84	757	0.813	ug/l #	79
27) cis-1,2-Dichloroethene	4.489	96	2542	2.508	ug/l	87
30) Chloroform	5.105	83	3227	1.825	ug/l	91
38) Carbon Tetrachloride	5.678	117	3601	2.723	ug/l #	87
44) Trichloroethene	7.129	130	352834	354.023	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	912	0.966	ug/l #	80
64) Tetrachloroethene	9.287	164	299	0.335	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
Data File : VX040465.D
Acq On : 26 Feb 2024 17:02
Operator : JC/MD
Sample : P1578-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20240222

Quant Time: Feb 27 05:11:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

