

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
 Data File : VX040468.D
 Acq On : 26 Feb 2024 18:11
 Operator : JC/MD
 Sample : P1578-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT119S1-20240220

Quant Time: Feb 27 05:12:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	65867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54373	49.355	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.720%
35) Dibromofluoromethane	5.391	113	40050	48.733	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.647	98	144979	48.600	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.079	95	59063	51.940	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.880%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	924	2.572	ug/l	96
44) Trichloroethene	7.129	130	2543	2.758	ug/l	96
64) Tetrachloroethene	9.281	164	573	0.707	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
Data File : VX040468.D
Acq On : 26 Feb 2024 18:11
Operator : JC/MD
Sample : P1578-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT119S1-20240220

Quant Time: Feb 27 05:12:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

