

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
 Data File : VX040471.D
 Acq On : 26 Feb 2024 19:21
 Operator : JC/MD
 Sample : P1578-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20240221

Quant Time: Feb 27 05:13:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	120089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56644	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56860	50.223	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.440%
35) Dibromofluoromethane	5.385	113	40079	47.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.980%
50) Toluene-d8	8.647	98	148069	48.847	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.700%
62) 4-Bromofluorobenzene	11.079	95	58878	50.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.900%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	1333	3.610	ug/l #	86
44) Trichloroethene	7.129	130	1696	1.810	ug/l	84
52) Toluene	8.726	92	1420	0.664	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022624\
Data File : VX040471.D
Acq On : 26 Feb 2024 19:21
Operator : JC/MD
Sample : P1578-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20240221

Quant Time: Feb 27 05:13:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

