

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022719\
 Data File : VX007740.D
 Acq On : 27 Feb 2019 18:04
 Operator : JC/SP
 Sample : K1594-31ME
 Misc : 6.68g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 28 03:59:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	255346	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	393479	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	382696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	205848	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	152605	48.25	ug/l	0.00
Spiked Amount			Recovery	=	96.50%	
35) Dibromofluoromethane	5.50	113	120829	45.94	ug/l	0.00
Spiked Amount			Recovery	=	91.88%	
50) Toluene-d8	8.71	98	489505	48.61	ug/l	0.00
Spiked Amount			Recovery	=	97.22%	
62) 4-Bromofluorobenzene	11.14	95	186155	50.37	ug/l	0.00
Spiked Amount			Recovery	=	100.74%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	1.64	94	586	Below Cal		84
18) Methyl Acetate	2.81	43	3969	1.293	ug/l #	62
20) Methylene Chloride	2.84	84	1446	0.557	ug/l #	88
25) 2-Butanone	4.74	43	827	0.386	ug/l #	65
31) Cyclohexane	5.55	56	3796	0.864	ug/l #	85
36) 1,1-Dichloropropene	5.65	75	17129	4.842	ug/l #	47
38) Carbon Tetrachloride	5.66	117	22407	6.306	ug/l #	17
39) Methylcyclohexane	7.46	83	2785	0.623	ug/l	94
53) t-1,3-Dichloropropene	8.97	75	425	2.897	ug/l #	86
60) Dibromochloromethane	9.33	129	822	0.297	ug/l #	12
64) Tetrachloroethene	9.34	164	968	0.312	ug/l #	85
67) Ethyl Benzene	10.25	91	91881	6.701	ug/l	100
68) m/p-Xylenes	10.36	106	77655	14.558	ug/l	97
73) Isopropylbenzene	11.02	105	6251	0.425	ug/l	97
76) 1,2,3-Trichloropropane	11.14	75	90907	20.718	ug/l #	32
78) n-propylbenzene	11.36	91	20860	1.255	ug/l	99
79) 2-Chlorotoluene	11.45	91	3071	0.307	ug/l #	38
80) 1,3,5-Trimethylbenzene	11.51	105	24714	2.042	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.14	75	90907	55.187	ug/l #	9
82) 4-Chlorotoluene	11.51	91	3037	0.259	ug/l #	42
83) tert-Butylbenzene	11.80	119	9121	0.785	ug/l #	64
84) 1,2,4-Trimethylbenzene	11.80	105	71127	5.746	ug/l	100
85) sec-Butylbenzene	11.80	105	71127	4.885	ug/l #	55
89) n-Butylbenzene	12.38	91	2314	0.203	ug/l #	27
90) Hexachloroethane	12.59	117	764	0.356	ug/l #	5
95) Naphthalene	13.83	128	26775	1.977	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022719\
Data File : VX007740.D
Acq On : 27 Feb 2019 18:04
Operator : JC/SP
Sample : K1594-31ME
Misc : 6.68g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :

Quant Time: Feb 28 03:59:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:168 Resp: 255346

Ion Ratio Lower Upper

168 100

99 49.3 37.9 56.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#5

Bromomethane

Concen: Below Cal

RT: 1.64 min Scan# 80

Delta R.T. 0.00 min

Lab File: VX007740.D


Acq: 27 Feb 2019 18:04

Tgt Ion: 94 Resp: 586

Ion Ratio Lower Upper

94 100

96 77.1 73.9 110.9



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

1.64

500

400

300


200


100

0

Time-->

1.60 1.65 1.70









#18
Methvl Acetate
Concen: 1.293 ug/l
RT: 2.81 min Scan# 271
Delta R.T. 0.02 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 3969
Ion Ratio Lower Upper
43 100
74 5.7 19.4 29.2#



#20
Methylene Chloride
Concen: 0.557 ug/l
RT: 2.84 min Scan# 277
Delta R.T. -0.02 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04

Tgt Ion: 84 Resp: 1446
Ion Ratio Lower Upper
84 100
49 129.5 94.8 142.2
51 34.8 29.8 44.8
86 51.4 54.1 81.1#









#25
2-Butanone
Concen: 0.386 ug/l
RT: 4.74 min Scan# 588
Delta R.T. 0.04 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 43 Resp: 827
Ion Ratio Lower Upper
43 100
72 45.9 22.2 33.2#



#31
Cyclohexane
Concen: 0.864 ug/l
RT: 5.55 min Scan# 721
Delta R.T. -0.02 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04

Tgt Ion: 56 Resp: 3796
Ion Ratio Lower Upper
56 100
69 21.8 23.0 34.6#
84 73.3 69.8 104.6





#33

1,2-Dichloroethane-d4

Concen: 48.248 ug/l

RT: 6.06 min Scan# 805

Delta R.T. -0.01 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 65 Resp: 152605

Ion Ratio Lower Upper

65 100

67 53.1 0.0 108.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04




Tgt Ion: 114 Resp: 393479


Ion Ratio Lower Upper

114 100

63 20.0 0.0 35.6

88 15.2 0.0 29.6





#35

Dibromofluoromethane

Concen: 45.938 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.00 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :
MSVOA_X
ClientSampled :


Tot Ion:113 Resp: 120829

Ion Ratio Lower Upper




113 100

111 102.6 81.8 122.8

192 24.3 18.5 27.7



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.842 ug/l

RT: 5.65 min Scan# 738

Delta R.T. -0.15 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04


Tgt Ion: 75 Resp: 17129

Ion Ratio Lower Upper



75 100


110 9.0 20.0 59.9#

77 0.0 24.7 37.1#



Abundance Ion 74.90 (74.60 to 75.60): VX0
Ion 109.90 (109.60 to 110.60): V
Ion 76.90 (76.60 to 77.60): VX0





#38

Carbon Tetrachloride

Concen: 6.306 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.13 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X

ClientSampleId :





Tot Ion:117 Resp: 22407

Ion Ratio Lower Upper

117 100

119 5.0 75.3 112.9#

121 0.0 25.0 37.4#



#39

Methylcyclohexane

Concen: 0.623 ug/l

RT: 7.46 min Scan# 1034

Delta R.T. -0.01 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04




Tgt Ion: 83 Resp: 2785


Ion Ratio Lower Upper

83 100

55 74.3 56.2 84.2

98 41.6 36.8 55.2









#50
Toluene-d8
Concen: 48.612 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 98 Resp: 489505
Ion Ratio Lower Upper
98 100
100 64.5 51.6 77.4



#53
t-1,3-Dichloropropene
Concen: 2.897 ug/l
RT: 8.97 min Scan# 1282
Delta R.T. -0.07 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04

Tgt Ion: 75 Resp: 425
Ion Ratio Lower Upper
75 100
77 23.9 25.2 37.8#





#60

Dibromochloromethane

Concen: 0.297 ug/l

RT: 9.33 min Scan# 1341

Delta R.T. -0.25 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X


ClientSampled :

Tot Ion:129 Resp: 822



Ion Ratio Lower Upper

129 100


127 0.0 37.8 113.4#



Abundance Ion 128.80 (128.50 to 129.50): V



Time-->



#62

4-Bromofluorobenzene

Concen: 50.367 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.01 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04


Tgt Ion: 95 Resp: 186155

Ion Ratio Lower Upper

95 100

174 91.8 0.0 189.2



176 89.0 0.0 186.2




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V




Time-->






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.01 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
117	100		
82	52.8	42.1	63.1
119	32.1	26.5	39.7






Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V




#64
Tetrachloroethene
Concen: 0.312 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. 0.00 min
Lab File: VX007740.D
Acq: 27 Feb 2019 18:04

Tgt Ion	Ratio	Lower	Upper
164	100		
166	108.8	101.0	151.4
129	66.4	70.5	105.7#
131	80.1	69.8	104.6

Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V





#67

Ethyl Benzene

Concen: 6.701 ug/l

RT: 10.25 min Scan# 1492

Delta R.T. 0.00 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X





ClientSampled :

Tot Ion: 91 Resp: 91881

Ion Ratio Lower Upper

91 100

106 31.4 25.4 38.0



#68

m/p-Xylenes

Concen: 14.558 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. 0.00 min

Lab File: VX007740.D




Acq: 27 Feb 2019 18:04


Tgt Ion: 106 Resp: 77655

Ion Ratio Lower Upper

106 100

91 199.9 156.6 234.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:152 Resp: 205848

Ion Ratio Lower Upper

152 100

115 55.3 38.0 114.0




150 157.5 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 0.425 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. 0.00 min

Lab File: VX007740.D


Acq: 27 Feb 2019 18:04

Tgt Ion:105 Resp: 6251

Ion Ratio Lower Upper



105 100


120 28.7 13.6 40.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#76

1,2,3-Trichloropropane

Concen: 20.718 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.16 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 75 Resp: 90907

Ion Ratio Lower Upper

75 100

77 1.2 23.0 69.0#



#78

n-propylbenzene

Concen: 1.255 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. 0.00 min

Lab File: VX007740.D




Acq: 27 Feb 2019 18:04


Tgt Ion: 91 Resp: 20860

Ion Ratio Lower Upper

91 100

120 24.4 12.4 37.2









#79
 2-Chlorotoluene
 Concen: 0.307 ug/l
 RT: 11.45 min Scan# 1689
 Delta R.T. 0.03 min
 Lab File: VX007740.D
 Acq: 27 Feb 2019 18:04




Instrument :
 MSVOA_X
 ClientSampled :


Tot Ion: 91 Resp: 3071
 Ion Ratio Lower Upper
 91 100
 126 0.0 18.7 56.1#



#80
 1,3,5-Trimethylbenzene
 Concen: 2.042 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. 0.00 min
 Lab File: VX007740.D
 Acq: 27 Feb 2019 18:04

Tgt Ion: 105 Resp: 24714
 Ion Ratio Lower Upper
 105 100
 120 49.9 25.3 75.8





#81

trans-1,4-Dichloro-2-butene

Concen: 55.187 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.07 min

Lab File: VX007740.D


Acq: 27 Feb 2019 18:04

Instrument :




MSVOA_X

ClientSampleId :

Tot Ion:	75	Resp:	90907
Ion	Ratio	Lower	Upper
75	100		
53	0.0	81.0	121.6#
89	0.0	41.4	62.0#



Abundance Ion 74.90 (74.60 to 75.60): VX0
Ion 53.00 (52.70 to 53.70): VX0
Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 0.259 ug/l


RT: 11.51 min Scan# 1698

Delta R.T. -0.01 min



Lab File: VX007740.D


Acq: 27 Feb 2019 18:04

Tgt Ion:	91	Resp:	3037
Ion	Ratio	Lower	Upper
91	100		
126	0.0	16.2	48.5#



Abundance Ion 91.00 (90.70 to 91.70): VX0
Ion 125.90 (125.60 to 126.60): V





#83

tert-Butylbenzene

Concen: 0.785 ug/l

RT: 11.80 min Scan# 1747

Delta R.T. 0.04 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:119 Resp: 9121

Ion Ratio Lower Upper

119 100

91 80.2 28.5 85.6




134 0.9 12.0 36.0#



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 91.00 (90.70 to 91.70): VX0

Ion 134.00 (133.70 to 134.70): V



#84

1,2,4-Trimethylbenzene

Concen: 5.746 ug/l

RT: 11.80 min Scan# 1747

Delta R.T. 0.00 min

Lab File: VX007740.D


Acq: 27 Feb 2019 18:04

Tgt Ion:105 Resp: 71127

Ion Ratio Lower Upper



105 100


120 46.0 23.1 69.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 4.885 ug/l

RT: 11.80 min Scan# 1747

Delta R.T. -0.14 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:105 Resp: 71127

Ion Ratio Lower Upper




105 100

134 0.1 10.6 31.8#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#89

n-Butylbenzene

Concen: 0.203 ug/l

RT: 12.38 min Scan# 1841

Delta R.T. -0.01 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04


Tgt Ion: 91 Resp: 2314

Ion Ratio Lower Upper

91 100

92 27.1 26.3 78.9



134 102.4 13.9 41.6#




Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 92.00 (91.70 to 92.70): VX0

Ion 134.00 (133.70 to 134.70): V





#90

Hexachloroethane

Concen: 0.356 ug/l

RT: 12.59 min Scan# 1876

Delta R.T. -0.01 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Instrument :

MSVOA_X





ClientSampled :

Tot Ion:117 Resp: 764

Ion Ratio Lower Upper

117 100

201 0.0 44.9 134.7#



#95

Naphthalene

Concen: 1.977 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. 0.00 min

Lab File: VX007740.D

Acq: 27 Feb 2019 18:04

Tgt Ion:128 Resp: 26775

Ion Ratio Lower Upper

128 100

127 13.9 10.5 15.7

129 11.1 8.7 13.1

