

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022719\
 Data File : VX007736.D
 Acq On : 27 Feb 2019 16:18
 Operator : JC/SP
 Sample : K1594-38DL 10X
 Misc : 6.24g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-8(7.5-8.5)DL

Quant Time: Feb 28 04:26:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	321314	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	499960	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	475818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	230318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	195073	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
35) Dibromofluoromethane	5.51	113	161105	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
50) Toluene-d8	8.71	98	614015	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
62) 4-Bromofluorobenzene	11.13	95	222154	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
Target Compounds						
67) Ethyl Benzene	10.25	91	4335	0.254	ug/l	100
73) Isopropylbenzene	11.02	105	3425	0.208	ug/l	100
78) n-propylbenzene	11.36	91	13644	0.733	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	3517	0.260	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	19406	1.401	ug/l	100
95) Naphthalene	13.83	128	6785	0.448	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022719\
Data File : VX007736.D
Acq On : 27 Feb 2019 16:18
Operator : JC/SP
Sample : K1594-38DL 10X
Misc : 6.24g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
B-8(7.5-8.5)DL

Quant Time: Feb 28 04:26:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

