

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022720\
 Data File : VX015040.D
 Acq On : 27 Feb 2020 12:13
 Operator : JC/SP
 Sample : L1683-14DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140N-20200225DL

Quant Time: Feb 27 16:08:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	429644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	687827	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	658747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	330825	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	244723	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
35) Dibromofluoromethane	5.49	113	207632	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	8.71	98	823913	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.14	95	303217	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	60785	12.514	ug/l	91
44) Trichloroethene	7.21	130	301269	56.452	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022720\
Data File : VX015040.D
Acq On : 27 Feb 2020 12:13
Operator : JC/SP
Sample : L1683-14DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140N-20200225DL

Quant Time: Feb 27 16:08:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

