

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022720\
 Data File : VX015045.D
 Acq On : 27 Feb 2020 14:06
 Operator : JC/SP
 Sample : L1683-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPRING-20200225

Quant Time: Feb 27 16:19:58 2020
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	378437	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	595930	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	576777	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	304256	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	204000	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
35) Dibromofluoromethane	5.49	113	180227	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
50) Toluene-d8	8.71	98	725546	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	11.13	95	272114	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
Target Compounds						
44) Trichloroethene	7.21	130	3200	0.692	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022720\
Data File : VX015045.D
Acq On : 27 Feb 2020 14:06
Operator : JC/SP
Sample : L1683-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPRING-20200225

Quant Time: Feb 27 16:19:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

