

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022723\
 Data File : VX034276.D
 Acq On : 27 Feb 2023 14:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 28 04:08:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	301843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	497000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	442287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	199318	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	175673	43.215	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.420%
35) Dibromofluoromethane	5.379	113	156698	48.424	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.647	98	592236	50.975	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	204395	45.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022723\
Data File : VX034276.D
Acq On : 27 Feb 2023 14:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 28 04:08:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

