



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022723\
Data File : VX034270.D
Acq On : 27 Feb 2023 08:14
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Title : SW846 8260
Last Update : Fri Feb 17 12:33:35 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	3958	PASS
75	95	30	60	46.9	9480	PASS
95	95	100	100	100.0	20205	PASS
96	95	5	9	6.5	1310	PASS
173	174	0.00	2	0.6	99	PASS
174	95	50	100	75.9	15344	PASS
175	174	5	9	7.9	1210	PASS
176	174	95	101	96.8	14854	PASS
177	176	5	9	6.8	1006	PASS