

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022724\
 Data File : VX040481.D
 Acq On : 27 Feb 2024 11:58
 Operator : JC/MD
 Sample : P1594-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB07-20240222

Quant Time: Feb 28 02:01:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	131664	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62163	49.519	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.040%
35) Dibromofluoromethane	5.385	113	44388	48.481	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.960%
50) Toluene-d8	8.647	98	164046	49.360	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.720%
62) 4-Bromofluorobenzene	11.079	95	63323	49.984	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022724\
Data File : VX040481.D
Acq On : 27 Feb 2024 11:58
Operator : JC/MD
Sample : P1594-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB07-20240222

Quant Time: Feb 28 02:01:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

