

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022724\
 Data File : VX040488.D
 Acq On : 27 Feb 2024 14:40
 Operator : JC/MD
 Sample : P1582-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 28 02:03:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116466	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57645	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54490	49.954	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.900%
35) Dibromofluoromethane	5.385	113	39090	48.265	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.540%
50) Toluene-d8	8.647	98	144274	49.076	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	56325	50.262	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022724\
Data File : VX040488.D
Acq On : 27 Feb 2024 14:40
Operator : JC/MD
Sample : P1582-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 28 02:03:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

