

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022724\
 Data File : VX040494.D
 Acq On : 27 Feb 2024 16:59
 Operator : JC/MD
 Sample : P1613-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1847

Quant Time: Feb 28 02:06:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	68689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118988	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53900	46.916	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.840%
35) Dibromofluoromethane	5.391	113	39783	48.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.160%
50) Toluene-d8	8.647	98	144078	47.970	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.940%
62) 4-Bromofluorobenzene	11.079	95	56232	49.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.240%
Target Compounds						
16) Acetone	2.380	43	1226	3.272	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022724\
Data File : VX040494.D
Acq On : 27 Feb 2024 16:59
Operator : JC/MD
Sample : P1613-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1847

Quant Time: Feb 28 02:06:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

