

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
 Data File : VX045062.D
 Acq On : 27 Feb 2025 14:55
 Operator : JC/MD
 Sample : Q1403-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Feb 28 00:43:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	38611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	103235	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	100796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	38343	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	40714	72.035	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	144.080%#
35) Dibromofluoromethane	5.379	113	36825	54.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.720%
50) Toluene-d8	8.647	98	115112	45.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.700%
62) 4-Bromofluorobenzene	11.079	95	45952	53.705	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
Data File : VX045062.D
Acq On : 27 Feb 2025 14:55
Operator : JC/MD
Sample : Q1403-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Feb 28 00:43:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

