

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
 Data File : VX045063.D
 Acq On : 27 Feb 2025 15:19
 Operator : JC/MD
 Sample : Q1403-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Feb 28 00:43:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	42372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	113910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43071	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	43197	69.645	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	139.280%#
35) Dibromofluoromethane	5.385	113	39978	53.980	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.960%
50) Toluene-d8	8.647	98	124872	44.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.180%
62) 4-Bromofluorobenzene	11.079	95	49510	52.441	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
Data File : VX045063.D
Acq On : 27 Feb 2025 15:19
Operator : JC/MD
Sample : Q1403-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Feb 28 00:43:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

