

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
 Data File : VX045064.D
 Acq On : 27 Feb 2025 15:42
 Operator : JC/MD
 Sample : Q1403-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Feb 28 00:44:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	42252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	109169	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	41094	66.442	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 132.880%#	
35) Dibromofluoromethane	5.385	113	37327	52.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.180%	
50) Toluene-d8	8.647	98	116925	43.564	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 87.120%	
62) 4-Bromofluorobenzene	11.079	95	47580	52.585	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
Data File : VX045064.D
Acq On : 27 Feb 2025 15:42
Operator : JC/MD
Sample : Q1403-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Feb 28 00:44:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

