

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
 Data File : VX045066.D
 Acq On : 27 Feb 2025 16:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 00:45:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.701	0.602	14.1	67	0.00
3 P	Chloromethane	0.854	0.667	21.9	62	0.00
4 C	Vinyl Chloride	0.827	0.697	15.7#	67	0.00
5 T	Bromomethane	0.247	0.267	-8.1	84	0.00
6 T	Chloroethane	0.307	0.326	-6.2	75	0.00
7 T	Trichlorofluoromethane	1.046	0.996	4.8	75	0.00
8 T	Diethyl Ether	0.379	0.368	2.9	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.579	8.4	72	0.00
10 T	Methyl Iodide	0.819	0.690	15.8	64	0.00
11 T	Tert butyl alcohol	0.136	0.141	-3.7	81	0.00
12 CM	1,1-Dichloroethene	0.644	0.569	11.6#	70	0.00
13 T	Acrolein	0.140	0.149	-6.4	82	0.00
14 T	Allyl chloride	1.165	1.077	7.6	73	0.00
15 T	Acrylonitrile	0.363	0.401	-10.5	82	0.00
16 T	Acetone	0.294	0.341	-16.0	91	0.00
17 T	Carbon Disulfide	1.767	1.317	25.5#	58	0.00
18 T	Methyl Acetate	0.928	1.197	-29.0#	99	0.00
19 T	Methyl tert-butyl Ether	2.050	2.013	1.8	77	0.00
20 T	Methylene Chloride	0.721	0.675	6.4	75	0.00
21 T	trans-1,2-Dichloroethene	0.634	0.572	9.8	70	0.00
22 T	Diisopropyl ether	2.203	2.218	-0.7	77	0.00
23 T	Vinyl Acetate	1.796	1.813	-0.9	76	0.00
24 P	1,1-Dichloroethane	1.233	1.199	2.8	76	0.00
25 T	2-Butanone	0.478	0.569	-19.0	88	0.00
26 T	2,2-Dichloropropane	0.978	0.798	18.4	65	0.00
27 T	cis-1,2-Dichloroethene	0.762	0.732	3.9	75	0.00
28 T	Bromochloromethane	0.593	0.598	-0.8	80	0.00
29 T	Tetrahydrofuran	0.317	0.370	-16.7	88	0.00
30 C	Chloroform	1.196	1.205	-0.8#	80	0.00
31 T	Cyclohexane	1.137	0.956	15.9	66	0.00
32 T	1,1,1-Trichloroethane	1.014	0.980	3.4	76	0.00
33 S	1,2-Dichloroethane-d4	0.732	0.774	-5.7	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.325	0.345	-6.2	83	0.00
36 T	1,1-Dichloropropene	0.469	0.425	9.4	70	0.00
37 T	Ethyl Acetate	0.522	0.584	-11.9	85	0.00
38 T	Carbon Tetrachloride	0.460	0.451	2.0	76	0.00
39 T	Methylcyclohexane	0.606	0.538	11.2	66	0.00
40 TM	Benzene	1.465	1.403	4.2	73	0.00
41 T	Methacrylonitrile	0.279	0.339	-21.5	86	0.00
42 TM	1,2-Dichloroethane	0.467	0.506	-8.4	82	0.00
43 T	Isopropyl Acetate	0.842	0.903	-7.2	79	0.00
44 TM	Trichloroethene	0.335	0.323	3.6	74	0.00
45 C	1,2-Dichloropropane	0.364	0.367	-0.8#	77	0.00
46 T	Dibromomethane	0.253	0.264	-4.3	79	0.00
47 T	Bromodichloromethane	0.489	0.525	-7.4	81	0.00
48 T	Methyl methacrylate	0.397	0.451	-13.6	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
 Data File : VX045066.D
 Acq On : 27 Feb 2025 16:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 00:45:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	82	0.00
50 S	Toluene-d8	1.229	1.243	-1.1	77	0.00
51 T	4-Methyl-2-Pentanone	0.512	0.612	-19.5	85	0.00
52 CM	Toluene	0.872	0.858	1.6#	73	0.00
53 T	t-1,3-Dichloropropene	0.495	0.476	3.8	71	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.540	2.2	72	0.00
55 T	1,1,2-Trichloroethane	0.335	0.346	-3.3	78	0.00
56 T	Ethyl methacrylate	0.542	0.558	-3.0	74	0.00
57 T	1,3-Dichloropropane	0.587	0.604	-2.9	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.284	-6.8	76	0.00
59 T	2-Hexanone	0.369	0.450	-22.0	85	0.00
60 T	Dibromochloromethane	0.359	0.380	-5.8	78	0.00
61 T	1,2-Dibromoethane	0.340	0.357	-5.0	78	0.00
62 S	4-Bromofluorobenzene	0.414	0.428	-3.4	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.315	0.302	4.1	74	0.00
65 PM	Chlorobenzene	1.074	1.073	0.1	74	0.00
66 T	1,1,1,2-Tetrachloroethane	0.346	0.356	-2.9	76	0.00
67 C	Ethyl Benzene	1.895	1.885	0.5#	74	0.00
68 T	m/p-Xylenes	0.699	0.693	0.9	73	0.00
69 T	o-Xylene	0.701	0.687	2.0	74	0.00
70 T	Styrene	1.130	1.159	-2.6	73	0.00
71 P	Bromoform	0.258	0.283	-9.7	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
73 T	Isopropylbenzene	4.026	3.950	1.9	74	0.00
74 T	N-amyl acetate	1.822	1.932	-6.0	77	0.00
75 P	1,1,2,2-Tetrachloroethane	1.383	1.424	-3.0	79	0.00
76 T	1,2,3-Trichloropropane	1.109	1.129	-1.8	77	0.00
77 T	Bromobenzene	0.913	0.914	-0.1	76	0.00
78 T	n-propylbenzene	4.647	4.583	1.4	73	0.00
79 T	2-Chlorotoluene	2.856	2.737	4.2	73	0.00
80 T	1,3,5-Trimethylbenzene	3.267	3.211	1.7	73	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.378	12.3	66	0.00
82 T	4-Chlorotoluene	3.160	3.064	3.0	73	0.00
83 T	tert-Butylbenzene	3.305	3.274	0.9	75	0.00
84 T	1,2,4-Trimethylbenzene	3.225	3.237	-0.4	73	0.00
85 T	sec-Butylbenzene	4.093	4.052	1.0	74	0.00
86 T	p-Isopropyltoluene	3.291	3.304	-0.4	73	0.00
87 T	1,3-Dichlorobenzene	1.678	1.645	2.0	74	0.00
88 T	1,4-Dichlorobenzene	1.697	1.641	3.3	74	0.00
89 T	n-Butylbenzene	2.912	2.956	-1.5	72	0.00
90 T	Hexachloroethane	0.616	0.585	5.0	72	0.00
91 T	1,2-Dichlorobenzene	1.650	1.647	0.2	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.278	-10.3	81	0.00
93 T	1,2,4-Trichlorobenzene	0.998	0.982	1.6	73	0.00
94 T	Hexachlorobutadiene	0.396	0.387	2.3	77	0.00
95 T	Naphthalene	3.608	3.790	-5.0	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022725\
Data File : VX045066.D
Acq On : 27 Feb 2025 16:29
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 28 00:45:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	1.031	-2.5	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6