

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\
 Data File : VX007756.D
 Acq On : 28 Feb 2019 10:31
 Operator : JC/SP
 Sample : VX0228WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0228WBL01

Quant Time: Mar 01 01:46:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	314817	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	489876	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	457356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	215481	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	190798	48.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.86%	
35) Dibromofluoromethane	5.50	113	161309	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
50) Toluene-d8	8.71	98	597809	47.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.36%	
62) 4-Bromofluorobenzene	11.13	95	209496	45.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\
Data File : VX007756.D
Acq On : 28 Feb 2019 10:31
Operator : JC/SP
Sample : VX0228WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0228WBL01

Quant Time: Mar 01 01:46:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

