

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\
 Data File : VX007759.D
 Acq On : 28 Feb 2019 11:59
 Operator : JC/SP
 Sample : K1603-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T1-OZ(PRE)

Quant Time: Mar 01 12:16:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	311129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	486490	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	464092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188885	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
35) Dibromofluoromethane	5.50	113	161273	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
50) Toluene-d8	8.71	98	598737	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
62) 4-Bromofluorobenzene	11.13	95	214370	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
Target Compounds						
19) Methyl tert-butyl Ether	3.20	73	14919	1.574	ug/l	98
24) 1,1-Dichloroethane	3.70	63	7297	1.282	ug/l	97
29) Tetrahydrofuran	5.16	42	8099	4.706	ug/l	93
44) Trichloroethene	7.21	130	29040	7.819	ug/l	88
64) Tetrachloroethene	9.34	164	13659	3.625	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\
Data File : VX007759.D
Acq On : 28 Feb 2019 11:59
Operator : JC/SP
Sample : K1603-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T1-OZ(PRE)

Quant Time: Mar 01 12:16:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

