

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\
 Data File : VX007760.D
 Acq On : 28 Feb 2019 12:25
 Operator : JC/SP
 Sample : K1603-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T2-OZ(PRE)

Quant Time: Mar 01 01:59:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	488069	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	465571	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	221466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	190736	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
35) Dibromofluoromethane	5.51	113	160284	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
50) Toluene-d8	8.71	98	599130	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
62) 4-Bromofluorobenzene	11.13	95	216399	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	5578	3.232	ug/l	100
19) Methyl tert-butyl Ether	3.21	73	5091	0.533	ug/l	92
24) 1,1-Dichloroethane	3.70	63	6731	1.174	ug/l	98
29) Tetrahydrofuran	5.16	42	4615	2.663	ug/l	88
44) Trichloroethene	7.21	130	4289	1.151	ug/l	93
64) Tetrachloroethene	9.34	164	4046	1.070	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\
Data File : VX007760.D
Acq On : 28 Feb 2019 12:25
Operator : JC/SP
Sample : K1603-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T2-OZ(PRE)

Quant Time: Mar 01 01:59:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

