

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\
 Data File : VX007761.D
 Acq On : 28 Feb 2019 12:52
 Operator : JC/SP
 Sample : K1603-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T3-OZ(PRE)

Quant Time: Mar 01 12:34:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	320228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	500950	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	476724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	224743	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	192900	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	5.51	113	164350	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
50) Toluene-d8	8.71	98	617238	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
62) 4-Bromofluorobenzene	11.13	95	219100	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
Target Compounds						
16) Acetone	2.45	43	5586	3.167	ug/l	Qvalue 100
24) 1,1-Dichloroethane	3.70	63	4837	0.825	ug/l	98
25) 2-Butanone	4.70	43	3491	1.299	ug/l	95
29) Tetrahydrofuran	5.16	42	4221	2.383	ug/l	95
30) Chloroform	5.21	83	1676	0.267	ug/l	98
44) Trichloroethene	7.22	130	1836	0.480	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\
Data File : VX007761.D
Acq On : 28 Feb 2019 12:52
Operator : JC/SP
Sample : K1603-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T3-OZ(PRE)

Quant Time: Mar 01 12:34:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

