

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\
 Data File : VX007763.D
 Acq On : 28 Feb 2019 13:45
 Operator : JC/SP
 Sample : K1641-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-460-462

Quant Time: Mar 01 07:13:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	315270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	491756	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	467547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	225616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	190110	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
35) Dibromofluoromethane	5.51	113	161560	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	600731	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
62) 4-Bromofluorobenzene	11.13	95	219943	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
Target Compounds						
16) Acetone	2.45	43	37551	21.623	ug/l	96
25) 2-Butanone	4.70	43	14161	5.353	ug/l	94
64) Tetrachloroethene	9.34	164	5042	1.328	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\
Data File : VX007763.D
Acq On : 28 Feb 2019 13:45
Operator : JC/SP
Sample : K1641-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-460-462

Quant Time: Mar 01 07:13:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

