

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\  
 Data File : VX007768.D  
 Acq On : 28 Feb 2019 15:59  
 Operator : JC/SP  
 Sample : K1655-04  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FB-022719

Quant Time: Mar 01 02:53:23 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X022519W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Feb 26 01:53:13 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|--------|-----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 311913   | 50.00 | ug/l   | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 486788   | 50.00 | ug/l   | 0.00      |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 468725   | 50.00 | ug/l   | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 229461   | 50.00 | ug/l   | 0.00      |
| System Monitoring Compounds |       |      |          |       |        |           |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 186946   | 48.39 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 96.78% |           |
| 35) Dibromofluoromethane    | 5.51  | 113  | 159179   | 48.92 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 97.84% |           |
| 50) Toluene-d8              | 8.71  | 98   | 603040   | 48.41 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 96.82% |           |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 218574   | 47.80 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 95.60% |           |
| Target Compounds            |       |      |          |       |        |           |
| 16) Acetone                 | 2.45  | 43   | 6243     | 3.634 | ug/l   | Qvalue 92 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\  
Data File : VX007768.D  
Acq On : 28 Feb 2019 15:59  
Operator : JC/SP  
Sample : K1655-04  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampled :  
FB-022719

Quant Time: Mar 01 02:53:23 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X022519W.M  
Quant Title : SW846 8260  
QLast Update : Tue Feb 26 01:53:13 2019  
Response via : Initial Calibration

