

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022819\
 Data File : VX007772.D
 Acq On : 28 Feb 2019 17:45
 Operator : JC/SP
 Sample : K1641-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-440-442

Quant Time: Mar 01 03:04:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	318517	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	496744	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	469256	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222811	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	191771	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
35) Dibromofluoromethane	5.50	113	161736	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
50) Toluene-d8	8.71	98	610250	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
62) 4-Bromofluorobenzene	11.13	95	216280	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
Target Compounds						
16) Acetone	2.45	43	8585	4.893	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	5090	1.321	ug/l	98
44) Trichloroethene	7.21	130	7645	2.016	ug/l	100
64) Tetrachloroethene	9.34	164	16921	4.442	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022819\
Data File : VX007772.D
Acq On : 28 Feb 2019 17:45
Operator : JC/SP
Sample : K1641-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-440-442

Quant Time: Mar 01 03:04:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

