

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022822\
 Data File : VX027218.D
 Acq On : 28 Feb 2022 12:41
 Operator : JC/MD
 Sample : VX0228WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0228WBL01

Quant Time: Mar 01 00:42:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97633	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	161869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59114	50.414	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.820%	
35) Dibromofluoromethane	5.391	113	50938	50.697	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.400%	
50) Toluene-d8	8.653	98	184466	49.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.200%	
62) 4-Bromofluorobenzene	11.079	95	67057	46.500	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022822\
Data File : VX027218.D
Acq On : 28 Feb 2022 12:41
Operator : JC/MD
Sample : VX0228WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0228WBL01

Quant Time: Mar 01 00:42:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

