

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022822\
 Data File : VX027228.D
 Acq On : 28 Feb 2022 18:23
 Operator : JC/MD
 Sample : N1688-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-35

Quant Time: Mar 01 00:45:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	93168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	155087	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65896	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56561	50.548	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.100%		
35) Dibromofluoromethane	5.397	113	48240	50.112	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.220%		
50) Toluene-d8	8.653	98	181999	50.563	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.120%		
62) 4-Bromofluorobenzene	11.079	95	67820	49.086	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.180%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.965	59	30710	141.059	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	30566	9.821	ug/l	98
22) Diisopropyl ether	3.770	45	5984	1.922	ug/l #	76
31) Cyclohexane	5.483	56	3674	2.192	ug/l	98
39) Methylcyclohexane	7.385	83	692	0.396	ug/l	90
40) Benzene	6.044	78	351015	85.436	ug/l	100
52) Toluene	8.720	92	1770	0.673	ug/l	90
67) Ethyl Benzene	10.195	91	17289	3.419	ug/l	97
68) m/p-Xylenes	10.305	106	765	0.393	ug/l	92
73) Isopropylbenzene	10.963	105	7215	1.530	ug/l	98
78) n-propylbenzene	11.305	91	6737	1.257	ug/l	97
95) Naphthalene	13.780	128	3517	0.760	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022822\
Data File : VX027228.D
Acq On : 28 Feb 2022 18:23
Operator : JC/MD
Sample : N1688-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-35

Quant Time: Mar 01 00:45:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

