

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040506.D
 Acq On : 28 Feb 2024 12:58
 Operator : JC/MD
 Sample : P1614-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-SP302-A-20240227

Quant Time: Feb 29 03:48:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	73631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129506	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60666	49.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.520%
35) Dibromofluoromethane	5.391	113	43665	48.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.980%
50) Toluene-d8	8.646	98	158137	48.375	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.760%
62) 4-Bromofluorobenzene	11.079	95	61262	49.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.320%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.330	101	7047	8.434	ug/l	91
16) Acetone	2.379	43	892	2.221	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040506.D
Acq On : 28 Feb 2024 12:58
Operator : JC/MD
Sample : P1614-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-SP302-A-20240227

Quant Time: Feb 29 03:48:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

