

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040508.D
 Acq On : 28 Feb 2024 13:45
 Operator : JC/MD
 Sample : P1614-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-SP303-A-20240227

Quant Time: Feb 29 03:49:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	130900	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61088	50.068	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.140%		
35) Dibromofluoromethane	5.385	113	44474	48.858	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.720%		
50) Toluene-d8	8.647	98	159568	48.293	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.580%		
62) 4-Bromofluorobenzene	11.079	95	62473	49.601	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.200%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	3566	4.308	ug/l	94
30) Chloroform	5.105	83	511	0.283	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040508.D
Acq On : 28 Feb 2024 13:45
Operator : JC/MD
Sample : P1614-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-SP303-A-20240227

Quant Time: Feb 29 03:49:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

