

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040510.D
 Acq On : 28 Feb 2024 14:31
 Operator : JC/MD
 Sample : P1614-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-SP303-B-20240227

Quant Time: Feb 29 03:49:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	130350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61586	48.909	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.385	113	43962	48.499	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.647	98	161327	49.031	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.060%
62) 4-Bromofluorobenzene	11.079	95	63633	50.735	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.460%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	3486	4.080	ug/l	98
16) Acetone	2.386	43	879	2.141	ug/l #	77
30) Chloroform	5.092	83	448	0.240	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040510.D
Acq On : 28 Feb 2024 14:31
Operator : JC/MD
Sample : P1614-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-SP303-B-20240227

Quant Time: Feb 29 03:49:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

