

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040515.D
 Acq On : 28 Feb 2024 16:49
 Operator : JC/MD
 Sample : P1615-14 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20240227

Quant Time: Feb 29 03:52:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129349	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62499	50.665	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.320%		
35) Dibromofluoromethane	5.385	113	44880	49.895	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.800%		
50) Toluene-d8	8.646	98	160011	49.008	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.020%		
62) 4-Bromofluorobenzene	11.079	95	64021	51.439	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 102.880%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.330	101	2410	2.879	ug/l	97
12) 1,1-Dichloroethene	2.318	96	790	0.940	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	344	0.328	ug/l	70
44) Trichloroethene	7.122	130	135888	134.669	ug/l	97
52) Toluene	8.726	92	543	0.236	ug/l	90
64) Tetrachloroethene	9.274	164	338	0.381	ug/l #	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040515.D
Acq On : 28 Feb 2024 16:49
Operator : JC/MD
Sample : P1615-14 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20240227

Quant Time: Feb 29 03:52:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

