

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040518.D
 Acq On : 28 Feb 2024 17:58
 Operator : JC/MD
 Sample : P1615-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT172S1-20240226

Quant Time: Feb 29 03:53:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	131584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61574	48.838	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	97.680%	
35) Dibromofluoromethane	5.391	113	44228	48.335	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.680%	
50) Toluene-d8	8.647	98	163587	49.252	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.500%	
62) 4-Bromofluorobenzene	11.079	95	64994	51.334	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	102.660%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	990	1.157	ug/l	93
16) Acetone	2.373	43	930	2.262	ug/l #	86
30) Chloroform	5.111	83	795	0.425	ug/l	87
44) Trichloroethene	7.129	130	31967	31.142	ug/l	98
52) Toluene	8.720	92	4274	1.825	ug/l	93
64) Tetrachloroethene	9.275	164	1382	1.525	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040518.D
Acq On : 28 Feb 2024 17:58
Operator : JC/MD
Sample : P1615-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT172S1-20240226

Quant Time: Feb 29 03:53:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

