

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040519.D
 Acq On : 28 Feb 2024 18:21
 Operator : JC/MD
 Sample : P1615-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT150S1-20240226

Quant Time: Feb 29 03:53:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128930	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60376	47.875	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.740%		
35) Dibromofluoromethane	5.385	113	43997	49.073	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.140%		
50) Toluene-d8	8.647	98	162351	49.886	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.780%		
62) 4-Bromofluorobenzene	11.079	95	64927	52.337	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.680%		
Target Compounds					Qvalue	
16) Acetone	2.380	43	872	2.120	ug/l	94
30) Chloroform	5.105	83	883	0.472	ug/l #	76
44) Trichloroethene	7.129	130	11886	11.818	ug/l	86
52) Toluene	8.720	92	928	0.404	ug/l	98
64) Tetrachloroethene	9.275	164	1237	1.370	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040519.D
Acq On : 28 Feb 2024 18:21
Operator : JC/MD
Sample : P1615-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT150S1-20240226

Quant Time: Feb 29 03:53:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

