

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040521.D
 Acq On : 28 Feb 2024 19:07
 Operator : JC/MD
 Sample : P1615-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20240226

Quant Time: Feb 29 03:54:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	73507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63778	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60421	49.145	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.280%		
35) Dibromofluoromethane	5.385	113	44087	49.970	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.940%		
50) Toluene-d8	8.647	98	159672	49.858	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.720%		
62) 4-Bromofluorobenzene	11.079	95	63324	51.872	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.740%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	1110	1.213	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	19328	23.171	ug/l	97
12) 1,1-Dichloroethene	2.325	96	3003	3.586	ug/l	88
16) Acetone	2.380	43	1126	2.808	ug/l #	89
27) cis-1,2-Dichloroethene	4.507	96	3739	3.578	ug/l	90
30) Chloroform	5.093	83	3241	1.778	ug/l	94
38) Carbon Tetrachloride	5.678	117	2469	1.880	ug/l	87
44) Trichloroethene	7.129	130	372979	376.847	ug/l	99
52) Toluene	8.726	92	1685	0.746	ug/l	95
55) 1,1,2-Trichloroethane	9.159	97	991	1.057	ug/l #	75
64) Tetrachloroethene	9.281	164	454	0.527	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040521.D
Acq On : 28 Feb 2024 19:07
Operator : JC/MD
Sample : P1615-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20240226

Quant Time: Feb 29 03:54:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

