

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040522.D
 Acq On : 28 Feb 2024 19:29
 Operator : JC/MD
 Sample : P1615-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20240226

Quant Time: Feb 29 03:54:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	133173	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62859	49.329	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.660%
35) Dibromofluoromethane	5.391	113	45869	49.530	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	8.647	98	165439	49.215	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.079	95	68522	53.475	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.940%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	240	0.253	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.331	101	143116	165.534	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1187	1.367	ug/l #	70
16) Acetone	2.374	43	1088	2.618	ug/l	94
27) cis-1,2-Dichloroethene	4.495	96	1971	1.820	ug/l	86
44) Trichloroethene	7.123	130	93060	89.577	ug/l	93
52) Toluene	8.720	92	3139	1.324	ug/l	92
64) Tetrachloroethene	9.275	164	10850	12.076	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040522.D
Acq On : 28 Feb 2024 19:29
Operator : JC/MD
Sample : P1615-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20240226

Quant Time: Feb 29 03:54:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

