

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX022824\
 Data File : VX040525.D
 Acq On : 28 Feb 2024 20:38
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 29 03:55:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	45.100	9.8	87	0.00
3 P	Chloromethane	50.000	44.821	10.4	88	0.00
4 C	Vinyl Chloride	50.000	45.888	8.2#	90	0.00
5 T	Bromomethane	50.000	53.555	-7.1	93	0.00
6 T	Chloroethane	50.000	45.260	9.5	90	0.00
7 T	Trichlorofluoromethane	50.000	49.907	0.2	97	0.00
8 T	Diethyl Ether	50.000	48.532	2.9	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.722	6.6	91	0.00
10 T	Methyl Iodide	50.000	47.473	5.1	92	0.00
11 T	Tert butyl alcohol	250.000	262.165	-4.9	99	0.00
12 CM	1,1-Dichloroethene	50.000	44.247	11.5#	89	0.00
13 T	Acrolein	250.000	333.934	-33.6#	132	0.00
14 T	Allyl chloride	50.000	46.241	7.5	90	0.00
15 T	Acrylonitrile	250.000	241.010	3.6	93	0.00
16 T	Acetone	250.000	293.686	-17.5	111	0.00
17 T	Carbon Disulfide	50.000	40.578	18.8	82	0.00
18 T	Methyl Acetate	50.000	52.072	-4.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	48.524	3.0	93	0.00
20 T	Methylene Chloride	50.000	45.453	9.1	92	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.426	11.1	87	0.00
22 T	Diisopropyl ether	50.000	46.615	6.8	90	0.00
23 T	Vinyl Acetate	250.000	238.267	4.7	91	0.00
24 P	1,1-Dichloroethane	50.000	46.091	7.8	91	0.00
25 T	2-Butanone	250.000	262.790	-5.1	99	0.00
26 T	2,2-Dichloropropane	50.000	43.532	12.9	86	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.160	9.7	91	0.00
28 T	Bromochloromethane	50.000	48.080	3.8	86	0.00
29 T	Tetrahydrofuran	250.000	245.924	1.6	95	0.00
30 C	Chloroform	50.000	47.715	4.6#	92	0.00
31 T	Cyclohexane	50.000	44.010	12.0	85	0.00
32 T	1,1,1-Trichloroethane	50.000	49.803	0.4	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.328	1.3	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	50.507	-1.0	96	0.00
36 T	1,1-Dichloropropene	50.000	46.381	7.2	91	0.00
37 T	Ethyl Acetate	50.000	49.903	0.2	95	0.00
38 T	Carbon Tetrachloride	50.000	50.415	-0.8	93	0.00
39 T	Methylcyclohexane	50.000	46.314	7.4	86	0.00
40 TM	Benzene	50.000	47.091	5.8	90	0.00
41 T	Methacrylonitrile	50.000	49.398	1.2	95	0.00
42 TM	1,2-Dichloroethane	50.000	51.576	-3.2	97	0.00
43 T	Isopropyl Acetate	50.000	49.483	1.0	94	0.00
44 TM	Trichloroethene	50.000	46.294	7.4	90	0.00
45 C	1,2-Dichloropropane	50.000	48.368	3.3#	92	0.00
46 T	Dibromomethane	50.000	48.872	2.3	92	0.00
47 T	Bromodichloromethane	50.000	49.228	1.5	91	0.00
48 T	Methyl methacrylate	50.000	52.015	-4.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
 Data File : VX040525.D
 Acq On : 28 Feb 2024 20:38
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 29 03:55:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1096.202	-9.6	98	0.00
50 S	Toluene-d8	50.000	49.541	0.9	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	260.694	-4.3	96	0.00
52 CM	Toluene	50.000	49.191	1.6#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	49.971	0.1	90	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.430	3.1	91	0.00
55 T	1,1,2-Trichloroethane	50.000	50.883	-1.8	91	0.00
56 T	Ethyl methacrylate	50.000	51.318	-2.6	91	0.00
57 T	1,3-Dichloropropane	50.000	49.031	1.9	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	252.780	-1.1	89	0.00
59 T	2-Hexanone	250.000	270.231	-8.1	97	0.00
60 T	Dibromochloromethane	50.000	50.982	-2.0	94	0.00
61 T	1,2-Dibromoethane	50.000	50.336	-0.7	92	0.00
62 S	4-Bromofluorobenzene	50.000	52.776	-5.6	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	48.883	2.2	96	0.00
65 PM	Chlorobenzene	50.000	46.318	7.4	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.662	0.7	95	0.00
67 C	Ethyl Benzene	50.000	47.792	4.4#	92	0.00
68 T	m/p-Xylenes	100.000	93.966	6.0	91	0.00
69 T	o-Xylene	50.000	47.328	5.3	90	0.00
70 T	Styrene	50.000	49.000	2.0	93	0.00
71 P	Bromoform	50.000	50.512	-1.0	95	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	48.166	3.7	93	0.00
74 T	N-amyl acetate	50.000	49.141	1.7	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.393	5.2	94	0.00
76 T	1,2,3-Trichloropropane	50.000	48.927	2.1	93	0.00
77 T	Bromobenzene	50.000	46.974	6.1	90	0.00
78 T	n-propylbenzene	50.000	46.820	6.4	91	0.00
79 T	2-Chlorotoluene	50.000	46.416	7.2	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.456	3.1	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.857	8.3	92	0.00
82 T	4-Chlorotoluene	50.000	47.892	4.2	93	0.00
83 T	tert-Butylbenzene	50.000	48.209	3.6	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.854	2.3	93	0.00
85 T	sec-Butylbenzene	50.000	48.051	3.9	92	0.00
86 T	p-Isopropyltoluene	50.000	47.693	4.6	93	0.00
87 T	1,3-Dichlorobenzene	50.000	46.357	7.3	92	0.00
88 T	1,4-Dichlorobenzene	50.000	45.822	8.4	93	0.00
89 T	n-Butylbenzene	50.000	46.917	6.2	91	0.00
90 T	Hexachloroethane	50.000	48.931	2.1	93	0.00
91 T	1,2-Dichlorobenzene	50.000	47.487	5.0	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.870	-5.7	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.508	3.0	94	0.00
94 T	Hexachlorobutadiene	50.000	46.044	7.9	92	0.00
95 T	Naphthalene	50.000	49.454	1.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022824\
Data File : VX040525.D
Acq On : 28 Feb 2024 20:38
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 29 03:55:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.810	0.4	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6