

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022825\
 Data File : VX045085.D
 Acq On : 28 Feb 2025 13:47
 Operator : JC/MD
 Sample : Q1423-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-EB-20250221

Quant Time: Feb 28 14:23:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68971	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136036	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56190	51.218	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.440%	
35) Dibromofluoromethane	5.385	113	45886	50.444	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.880%	
50) Toluene-d8	8.647	98	169589	51.426	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	11.079	95	57093	52.246	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.500%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6489	12.748	ug/l	96
20) Methylene Chloride	2.788	84	643	0.638	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022825\
Data File : VX045085.D
Acq On : 28 Feb 2025 13:47
Operator : JC/MD
Sample : Q1423-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-EB-20250221

Quant Time: Feb 28 14:23:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

