

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
 Data File : VX040529.D
 Acq On : 29 Feb 2024 12:06
 Operator : JC/MD
 Sample : VX0229WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0229WBL01

Quant Time: Mar 01 03:44:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129380	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67637	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59183	47.268	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.540%
35) Dibromofluoromethane	5.385	113	44601	49.573	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.140%
50) Toluene-d8	8.647	98	160173	49.046	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.100%
62) 4-Bromofluorobenzene	11.079	95	66978	53.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
Data File : VX040529.D
Acq On : 29 Feb 2024 12:06
Operator : JC/MD
Sample : VX0229WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0229WBL01

Quant Time: Mar 01 03:44:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

