

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
 Data File : VX040533.D
 Acq On : 29 Feb 2024 13:41
 Operator : JC/MD
 Sample : P1615-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20240226DL

Quant Time: Mar 01 03:47:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	77829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	135145	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65100	50.010	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.020%	
35) Dibromofluoromethane	5.391	113	45918	48.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.720%	
50) Toluene-d8	8.646	98	157516	46.175	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 92.340%	
62) 4-Bromofluorobenzene	11.079	95	64037	49.245	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.500%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.330	101	1849	2.094	ug/l	91
44) Trichloroethene	7.122	130	34193	32.433	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
Data File : VX040533.D
Acq On : 29 Feb 2024 13:41
Operator : JC/MD
Sample : P1615-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20240226DL

Quant Time: Mar 01 03:47:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

