

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
 Data File : VX040527.D
 Acq On : 29 Feb 2024 10:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 01 03:43:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.622	0.594	4.5	98	0.00
3 P	Chloromethane	0.617	0.551	10.7	94	0.00
4 C	Vinyl Chloride	0.667	0.602	9.7#	95	0.00
5 T	Bromomethane	0.371	0.415	-11.9	104	0.00
6 T	Chloroethane	0.422	0.386	8.5	97	0.00
7 T	Trichlorofluoromethane	1.054	0.947	10.2	94	0.00
8 T	Diethyl Ether	0.368	0.308	16.3	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.561	1.1	103	0.00
10 T	Methyl Iodide	0.548	0.613	-11.9	108	0.00
11 T	Tert butyl alcohol	0.141	0.140	0.7	100	0.00
12 CM	1,1-Dichloroethene	0.570	0.507	11.1#	96	0.00
13 T	Acrolein	0.090	0.108	-20.0	126	0.00
14 T	Allyl chloride	0.997	0.946	5.1	99	0.00
15 T	Acrylonitrile	0.353	0.318	9.9	93	0.00
16 T	Acetone	0.273	0.346	-26.7#	128	0.00
17 T	Carbon Disulfide	1.578	1.287	18.4	88	0.00
18 T	Methyl Acetate	0.841	0.798	5.1	94	0.00
19 T	Methyl tert-butyl Ether	1.943	1.885	3.0	100	0.00
20 T	Methylene Chloride	0.653	0.585	10.4	96	0.00
21 T	trans-1,2-Dichloroethene	0.608	0.545	10.4	94	0.00
22 T	Diisopropyl ether	2.003	1.881	6.1	97	0.00
23 T	Vinyl Acetate	1.835	1.778	3.1	99	0.00
24 P	1,1-Dichloroethane	1.168	1.086	7.0	98	0.00
25 T	2-Butanone	0.475	0.491	-3.4	104	0.00
26 T	2,2-Dichloropropane	0.930	0.930	0.0	106	0.00
27 T	cis-1,2-Dichloroethene	0.711	0.657	7.6	99	0.00
28 T	Bromochloromethane	0.489	0.466	4.7	91	0.00
29 T	Tetrahydrofuran	0.331	0.302	8.8	95	0.00
30 C	Chloroform	1.240	1.192	3.9#	99	0.00
31 T	Cyclohexane	0.947	0.861	9.1	94	0.00
32 T	1,1,1-Trichloroethane	1.058	1.055	0.3	104	0.00
33 S	1,2-Dichloroethane-d4	0.836	0.818	2.2	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.348	0.358	-2.9	103	0.00
36 T	1,1-Dichloropropene	0.536	0.518	3.4	100	0.00
37 T	Ethyl Acetate	0.586	0.561	4.3	95	0.00
38 T	Carbon Tetrachloride	0.518	0.541	-4.4	102	0.00
39 T	Methylcyclohexane	0.578	0.581	-0.5	98	0.00
40 TM	Benzene	1.450	1.386	4.4	95	0.00
41 T	Methacrylonitrile	0.327	0.318	2.8	99	0.00
42 TM	1,2-Dichloroethane	0.600	0.637	-6.2	104	0.00
43 T	Isopropyl Acetate	0.920	0.916	0.4	99	0.00
44 TM	Trichloroethene	0.390	0.384	1.5	101	0.00
45 C	1,2-Dichloropropane	0.388	0.374	3.6#	97	0.00
46 T	Dibromomethane	0.285	0.282	1.1	98	0.00
47 T	Bromodichloromethane	0.562	0.567	-0.9	98	0.00
48 T	Methyl methacrylate	0.460	0.482	-4.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
 Data File : VX040527.D
 Acq On : 29 Feb 2024 10:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 01 03:43:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	101	0.00
50 S	Toluene-d8	1.262	1.276	-1.1	100	0.00
51 T	4-Methyl-2-Pentanone	0.608	0.619	-1.8	98	0.00
52 CM	Toluene	0.890	0.898	-0.9#	99	0.00
53 T	t-1,3-Dichloropropene	0.570	0.568	0.4	94	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.631	-2.9	101	0.00
55 T	1,1,2-Trichloroethane	0.369	0.353	4.3	89	0.00
56 T	Ethyl methacrylate	0.571	0.567	0.7	92	0.00
57 T	1,3-Dichloropropane	0.643	0.598	7.0	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.299	0.302	-1.0	93	0.00
59 T	2-Hexanone	0.468	0.456	2.6	92	0.00
60 T	Dibromochloromethane	0.401	0.397	1.0	96	0.00
61 T	1,2-Dibromoethane	0.403	0.391	3.0	93	0.00
62 S	4-Bromofluorobenzene	0.481	0.520	-8.1	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.351	0.353	-0.6	95	0.00
65 PM	Chlorobenzene	1.064	1.040	2.3	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.393	-4.8	96	0.00
67 C	Ethyl Benzene	1.947	2.000	-2.7#	94	0.00
68 T	m/p-Xylenes	0.727	0.750	-3.2	96	0.00
69 T	o-Xylene	0.695	0.768	-10.5	101	0.00
70 T	Styrene	1.159	1.293	-11.6	101	0.00
71 P	Bromoform	0.307	0.358	-16.6	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.682	3.669	0.4	102	0.00
74 T	N-amyl acetate	1.710	1.711	-0.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.332	1.249	6.2	98	0.00
76 T	1,2,3-Trichloropropane	1.100	1.108	-0.7	101	0.00
77 T	Bromobenzene	0.871	0.862	1.0	101	0.00
78 T	n-propylbenzene	4.567	4.551	0.4	102	0.00
79 T	2-Chlorotoluene	2.750	2.685	2.4	103	0.00
80 T	1,3,5-Trimethylbenzene	3.132	3.201	-2.2	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.349	0.343	1.7	101	0.00
82 T	4-Chlorotoluene	3.180	3.172	0.3	103	0.00
83 T	tert-Butylbenzene	2.952	3.040	-3.0	104	0.00
84 T	1,2,4-Trimethylbenzene	3.107	3.208	-3.3	104	0.00
85 T	sec-Butylbenzene	3.879	4.007	-3.3	105	0.00
86 T	p-Isopropyltoluene	3.197	3.314	-3.7	106	0.00
87 T	1,3-Dichlorobenzene	1.736	1.711	1.4	104	0.00
88 T	1,4-Dichlorobenzene	1.767	1.726	2.3	105	0.00
89 T	n-Butylbenzene	3.145	3.289	-4.6	107	0.00
90 T	Hexachloroethane	0.532	0.557	-4.7	106	0.00
91 T	1,2-Dichlorobenzene	1.654	1.656	-0.1	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.310	0.328	-5.8	105	0.00
93 T	1,2,4-Trichlorobenzene	1.079	1.136	-5.3	108	0.00
94 T	Hexachlorobutadiene	0.441	0.473	-7.3	113	0.00
95 T	Naphthalene	3.449	3.564	-3.3	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
Data File : VX040527.D
Acq On : 29 Feb 2024 10:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 01 03:43:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.026	1.081	-5.4	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6