

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
 Data File : VX040539.D
 Acq On : 29 Feb 2024 15:59
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 01 03:49:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62053	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59965	49.791	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.580%
35) Dibromofluoromethane	5.385	113	42320	49.217	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.440%
50) Toluene-d8	8.647	98	153141	49.065	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.079	95	62515	52.544	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.080%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022924\
Data File : VX040539.D
Acq On : 29 Feb 2024 15:59
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 01 03:49:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 713
Delta R.T. 0.006 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 72006
Ion Ratio Lower Upper
168 100
99 64.3 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 49.791 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.000 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59

Tgt Ion: 65 Resp: 59965
Ion Ratio Lower Upper
65 100
67 49.3 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.006 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 123651
Ion Ratio Lower Upper
114 100
63 22.2 0.0 45.0
88 16.1 0.0 36.6



#35
Dibromofluoromethane
Concen: 49.217 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59

Tgt Ion:113 Resp: 42320
Ion Ratio Lower Upper
113 100
111 104.3 82.9 124.3
192 17.3 13.3 19.9









#50
Toluene-d8
Concen: 49.065 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 153141
Ion Ratio Lower Upper
98 100
100 63.9 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 52.544 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59

Tgt Ion: 95 Resp: 62515
Ion Ratio Lower Upper
95 100
174 73.4 0.0 135.6
176 71.3 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 108084
Ion Ratio Lower Upper
117 100
82 57.1 46.2 69.4
119 29.5 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX040539.D
Acq: 29 Feb 2024 15:59

Tgt Ion:152 Resp: 62053
Ion Ratio Lower Upper
152 100
115 63.5 43.3 129.9
150 160.1 0.0 345.0

