

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030119\
 Data File : VX007785.D
 Acq On : 01 Mar 2019 13:13
 Operator : JC/SP
 Sample : K1655-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-022719

Quant Time: Mar 01 22:35:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	333864	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	518155	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	490236	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	232580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	191214	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
35) Dibromofluoromethane	5.51	113	168299	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
50) Toluene-d8	8.71	98	635496	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	11.13	95	228010	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
Target Compounds						
16) Acetone	2.45	43	9438	5.132	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030119\
Data File : VX007785.D
Acq On : 01 Mar 2019 13:13
Operator : JC/SP
Sample : K1655-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-022719

Quant Time: Mar 01 22:35:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

