

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030119\
 Data File : VX007795.D
 Acq On : 01 Mar 2019 17:40
 Operator : JC/SP
 Sample : K1701-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M2-T5-OZ(POST-HYPO)

Quant Time: Mar 01 23:12:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.61	168	95488	50.00	ug/l	-0.06
34) 1,4-Difluorobenzene	6.83	114	60148	50.00	ug/l	-0.03
63) Chlorobenzene-d5	10.11	117	173866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	140724	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	4639	3.92	ug/l	-0.04
Spiked Amount			Recovery	=	7.84%	
35) Dibromofluoromethane	5.45	113	3436	8.55	ug/l	-0.06
Spiked Amount			Recovery	=	17.10%	
50) Toluene-d8	8.70	98	189144	122.88	ug/l	-0.01
Spiked Amount			Recovery	=	245.76%	
62) 4-Bromofluorobenzene	11.13	95	112152	198.51	ug/l	0.00
Spiked Amount			Recovery	=	397.02%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030119\
Data File : VX007795.D
Acq On : 01 Mar 2019 17:40
Operator : JC/SP
Sample : K1701-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M2-T5-OZ(POST-HYPO)

Quant Time: Mar 01 23:12:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

