

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027243.D
 Acq On : 01 Mar 2022 13:09
 Operator : JC/MD
 Sample : N1688-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16R

Quant Time: Mar 02 00:34:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	152196	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54467	49.828	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.660%
35) Dibromofluoromethane	5.391	113	47533	50.315	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.640%
50) Toluene-d8	8.653	98	173814	49.206	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.420%
62) 4-Bromofluorobenzene	11.079	95	62918	46.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4145	8.637	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	14793	4.865	ug/l	98
40) Benzene	6.044	78	4552	1.129	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027243.D
Acq On : 01 Mar 2022 13:09
Operator : JC/MD
Sample : N1688-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16R

Quant Time: Mar 02 00:34:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

