

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027252.D
 Acq On : 01 Mar 2022 16:40
 Operator : JC/MD
 Sample : N1688-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-33

Quant Time: Mar 02 00:36:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	82577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	139869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49884	50.299	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.600%	
35) Dibromofluoromethane	5.391	113	43410	50.001	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.000%	
50) Toluene-d8	8.653	98	160610	49.475	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.960%	
62) 4-Bromofluorobenzene	11.079	95	61012	48.963	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.920%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1260	2.894	ug/l #	89
19) Methyl tert-butyl Ether	3.117	73	6162	2.234	ug/l #	10
22) Diisopropyl ether	3.770	45	2945	1.067	ug/l #	70
39) Methylcyclohexane	7.379	83	2905	1.843	ug/l	92
67) Ethyl Benzene	10.195	91	2750	0.604	ug/l	97
68) m/p-Xylenes	10.305	106	1015	0.580	ug/l	98
69) o-Xylene	10.646	106	792	0.451	ug/l	98
78) n-propylbenzene	11.305	91	5217	1.056	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	12440	3.405	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	20373	5.606	ug/l	98
85) sec-Butylbenzene	11.896	105	1539	0.346	ug/l #	56
86) p-Isopropyltoluene	12.012	119	1248	0.329	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027252.D
Acq On : 01 Mar 2022 16:40
Operator : JC/MD
Sample : N1688-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-33

Quant Time: Mar 02 00:36:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

