

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027253.D
 Acq On : 01 Mar 2022 17:03
 Operator : JC/MD
 Sample : N1688-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-34

Quant Time: Mar 02 00:36:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	84592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	141349	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50837	50.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.080%	
35) Dibromofluoromethane	5.391	113	44573	50.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.600%	
50) Toluene-d8	8.653	98	164104	50.022	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.040%	
62) 4-Bromofluorobenzene	11.079	95	62818	49.884	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.760%	
Target Compounds						Qvalue
11) Tert butyl alcohol	2.959	59	57180	289.269	ug/l	96
16) Acetone	2.386	43	1529	3.428	ug/l	94
19) Methyl tert-butyl Ether	3.123	73	60599	21.445	ug/l	100
22) Diisopropyl ether	3.770	45	10614	3.755	ug/l #	77
31) Cyclohexane	5.470	56	5657	3.718	ug/l	95
39) Methylcyclohexane	7.385	83	1542	0.968	ug/l #	86
40) Benzene	6.044	78	459397	122.684	ug/l	100
52) Toluene	8.720	92	1388	0.579	ug/l	99
67) Ethyl Benzene	10.195	91	28715	6.192	ug/l	98
68) m/p-Xylenes	10.305	106	1812	1.016	ug/l	92
73) Isopropylbenzene	10.963	105	10423	2.411	ug/l	100
78) n-propylbenzene	11.305	91	10023	2.040	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	12952	3.582	ug/l	100
85) sec-Butylbenzene	11.890	105	1714	0.387	ug/l #	56
95) Naphthalene	13.780	128	7981	1.880	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027253.D
Acq On : 01 Mar 2022 17:03
Operator : JC/MD
Sample : N1688-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-34

Quant Time: Mar 02 00:36:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

