

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027256.D
 Acq On : 01 Mar 2022 18:14
 Operator : JC/MD
 Sample : N1710-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HP-1-50

Quant Time: Mar 02 00:37:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	83893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	138361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63471	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50430	50.052	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.100%	
35) Dibromofluoromethane	5.391	113	43720	50.907	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.820%	
50) Toluene-d8	8.647	98	163969	51.060	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.120%	
62) 4-Bromofluorobenzene	11.079	95	60457	49.046	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.100%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2594	5.864	ug/l #	84
30) Chloroform	5.105	83	2744	1.667	ug/l	89
39) Methylcyclohexane	7.379	83	1984	1.272	ug/l	98
67) Ethyl Benzene	10.195	91	2572	0.565	ug/l	96
68) m/p-Xylenes	10.305	106	896	0.512	ug/l	99
73) Isopropylbenzene	10.963	105	4519	0.995	ug/l	96
78) n-propylbenzene	11.305	91	11360	2.201	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	3767	0.986	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	24045	6.330	ug/l	98
85) sec-Butylbenzene	11.890	105	8318	1.786	ug/l	86
86) p-Isopropyltoluene	12.012	119	4005	1.009	ug/l	96
89) n-Butylbenzene	12.335	91	7479	2.254	ug/l #	55
95) Naphthalene	13.780	128	41696	9.349	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027256.D
Acq On : 01 Mar 2022 18:14
Operator : JC/MD
Sample : N1710-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HP-1-50

Quant Time: Mar 02 00:37:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

