

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027258.D
 Acq On : 01 Mar 2022 19:01
 Operator : JC/MD
 Sample : N1725-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-02

Quant Time: Mar 02 00:37:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	78897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	132951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48706	51.402	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.800%
35) Dibromofluoromethane	5.391	113	41741	50.580	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	8.653	98	153298	49.680	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	57092	48.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.400%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1800	4.327	ug/l	98
20) Methylene Chloride	2.794	84	367	0.426	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027258.D
Acq On : 01 Mar 2022 19:01
Operator : JC/MD
Sample : N1725-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-02

Quant Time: Mar 02 00:37:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

